

As Left Certificate of Calibration

Manufacturer Radian Research, Inc.
Instrument Model: RD-33-333 Dytronic Portable Standard
Serial Number 300273
Firmware Revision: 07.34.06
Error Specification .01% worst case

ISO 9001:2015 Certified Quality Management System
 ISO/IEC:17025 Accredited Laboratory

Customer Name: Enercom, S.A.
Address: Calle Akee No. 1511-B
 Apdo. 0843-00289; Balboa, Ancon

Calibration Lab Name: Radian Research, Inc.
Calibration Lab Address: 3852 Fortune Drive; Lafayette, IN 47905
Calibration Date: 29-Oct-19
Report Issue Date: 14-Nov-19

Environmental Conditions

Temperature: 23°C +/- 2°C
Humidity: between 30% and 60%
P.O.Number: E-1988 / Wire Transfer
RMA / Certificate Number: RO37237

Physical Condition: Undamaged

Radian Research's As-Found Test Results showed this Instrument to be:
 New In Tolerance Out of Tolerance Inoperative Limited Calibration Repair Needed
Radian Research's As-Left Test Results showed this Instrument to be:
 New In Tolerance Out of Tolerance Inoperative Limited Calibration
 For Out of Tolerance conditions, As-Found Data Reports are furnished.

Radian Research, Inc. certifies the instrument listed above meets or exceeds specifications and was calibrated in compliance with ANSI/NC SL Z540-1-1994 "Calibration Laboratories and Measuring and Test Equipment – General Requirements" using applicable Radian Research procedures which meet the requirements of ISO 9001:2015 and ISO/IEC Guide 98-3 "Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)".


This instrument was calibrated by a Radian Research RS-933 Syntron Automated Calibration System which is traceable to the National Institute of Standards and Technology (NIST) and/or other National Metrology Institute (NMI). The Automated Calibration system is cross checked and calibrated on a schedule which is adjusted to maintain required accuracies and traceability. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2.

Procedure used for Calibration: 9912164
Software used for Calibration: Radian Research RS-933 Control Application Version 2.3.7142.30113
RS-933 Tool Number: 9911644

Standards Used in This Calibration:

<u>Description</u>	<u>Serial Number(s):</u>	<u>Cal. Source</u>	<u>Cal. Number:</u>	<u>Cal. Due Date:</u>
RS-940A Data Collection Module	933073	Radian Research	933073-20190330	3/30/2020
RS-711 Syntron with RS-932 Boost Module (Phase A)	703236	Radian Research	703236-20190508	5/8/2020
RS-711 Syntron with RS-932 Boost Module (Phase B)	703194	Radian Research	703194-20190423	4/23/2020
RS-711 Syntron with RS-932 Boost Module (Phase C)	703226	Radian Research	703226-20190327	3/27/2020

Laboratory Technician Signature



*Template 9903256.05
 This report shall not be reproduced, except in full, without prior written approval of the Calibration Facility*